

Code 3023-G

25 Test Strips

Ammonia
Fresh and Salt Water



Insta-Test[®] ANALYTIC

NH₃ • Ammonia ppm • Fresh and Salt Water




Immerse strip for 5 seconds.

Remove with **pad face up**.

**DO NOT SHAKE OFF
EXCESS WATER.**

Wait 60 seconds.

Compare to color chart.



LaMotte Company
Chester town, Maryland 21620 USA
www.lamotte.com

View from
above.



- Keep wet fingers out of the vial.
- Close vial tightly.
- Store in a cool, dry place.



6 37385 46305 1

Drop holding box and dotted tolerance line