



Wealtec Corp.

HB-1 / HB-2 Block Heater

Installation and Operation Instruction Manual

Item # 03040

Version 1.30

**This instrument is intended for laboratory use only*

EDW07J1005 -V2

<http://www.wealtec.com>

E-mail : support@wealtec.com

Packing List

- ☒ x 1 main body
- ☒ x 1 power cable
- ☒ x 1 manual (packing list & warranty card)
- ☒ x 1 block extractor

WEALTEC

Wealtec Corp.

1885 Meadowvale Way Sparks, NV 89431 Tel:+1-775-351-2066 Fax:+1-775-351-2077
<http://www.wealtec.com> E-mail: support@wealtec.com

Index

A. Important Notice	2
Warranty	
Technical and Service contact	
Safety notice	
Certification	
Caution Information	
B. Introduction	5
Specifications	
Product Description	
C. Installation of Block Heater	11
Package list	
Optional items	
Installation	
D. Operation	12
E. Care and maintenance	13
F. OFFSET function	14
G. Order Information	17

A. Important Notice

The instrument is designed for laboratory usage only.

Before setting up and operating Block Heater, please carefully read these instructions to get familiarized with product's installation and operation process. Instructions shall be read by well-trained individuals or by technical people from Wealtec Corp. before operating the instruments.

Any improper usage of the instruments may cause damage, please refer to the safety notice.

The instrument shall not be modified or altered in any way. Any modification or alteration will void the warranty, void the regulatory certifications and create potential safety hazard. Wealtec is not responsible for any injury or damage caused by using the instrument for any non-intended purpose or by modifying the instrument by any person who is not authorized by Wealtec Corp.

Warranty

Block Heater is warranted to be free from defects in materials and workmanship for a period of one year from the original invoice date under normal usage. Any defects to occur during warranty period, Wealtec Corp. will repair or replace defective products or parts without charge unless the defects arise from below conditions. The defects described below are specially excluded from Wealtec warranty policy.

1. Improper operation of the instrument
2. Repair or modification by any person who is not authorized by Wealtec Corp.
3. Damage caused by any (in)-direct accident, neglect or misuse.
4. Damage caused by disaster.
5. Damage caused by any improper solvents or samples

Technical and Service contact

Most of the operation details are described in the instruction manual. Operator can find the possible solution according to the instructions. For any other technical/ service questions, please contact your local representative or contact Wealtec world-wide service center by E-mail: support@wealtec.com.

Safety notice

Certification

Wealtec Block Heater is designed to meet the international electrical safety standard EN61010 and EMC regulations. This product is with CE marking. Operation according to the guidance of instruction manual is certificated safe. Any modification or alteration will void the warranty, void the regulatory certifications and create potential safety hazard.



Caution Information

1. Apply 10% higher or lower voltage than specified on the instrument may cause malfunction of the Block Heater.
2. Do not splash water or liquid on the block heater, the spilled water or liquid may cause malfunction of the Block Heater.
3. Do not touch or exchange the block while heating or when block is still hot, it may scald the operator.

4. Operate the Block Heater with volatile liquid may cause damage while heating the liquid.
5. The lifetime of the electronic components may be shorten if the Block Heater is operated with volatile liquid in a poor ventilated environment.
6. Do not directly pour water/liquid/ice on high temperature heating block to cool down the temperature of the block. It may cause damage to the instrument and fast-evaporated water/ liquid may be harmful to the operator.

B. Introduction

Accurate chemical reaction needs to have well-controlled temperature incubation. For fast heat transfer, Wealtec block heater offer three types of interchangeable blocks specified for test tube, micro-centrifuge tube and PCR reaction tubes. Broad range temperature control from ambient to 200°C is ideal for the applications of restriction digestion, denaturing of nucleic acids & proteins, hot start PCR, enzyme kinetics reactions, immunoassays , COD reactions, BUN, coagulation studies, melting agar etc.

Specifications

Module Type	HB-1	HB-2
Temperature Range	30 to 200°C	30 to 200°C
Temperature stability @ 100°C	+/- 0.1°C	+/- 0.1°C
Temperature uniformity @ 100°C	+/- 0.1°C	+/- 0.1°C
Temperature accuracy @ 100°C	+/- 0.1°C	+/- 0.1°C
Heat up speed from ambient to 100°C	15 +/- 2 minutes	15 +/- 2 minutes
Temperature controller	Microprocessor	Microprocessor
Temperature control mode	Digital	Digital
Temperature Calibration	Yes (front panel control)	Yes (front panel control)
Heat power (w)	250 W max	500 W max
Over Temp cutoff	212°C	212°C
Number of blocks	1	2
Block material	Solid anodized aluminum	Solid anodized aluminum
Bath internal dimensions	102 x 78 x 46mm (L x W x H)	156 x 102 x 46 mm (L x W x H)
Timer	0-999 minutes	0-999 minutes
Panel Display mode	2 x 3 digital display	2 x 3 digital display
Power	120/230 V	120/230 V

Dimensions	310 x 250 x 80 mm (L x W x H)	310 x 250 x 80 mm (L x W x H)
Weight with block	4.5 kg	6 kg
Weight without block	3.3 kg	4.8 kg
Warranty	1 year	1 year
Certification	CE	CE
Operating condition	Temperature: 0-40°C Humidity: 10% to 90% R.H. Non-condensing	
Block type	12 x 16 mm test tubes, 20 x 1.5 ml microtubes, 30 x 0.5 ml microtubes	


Product Description


Wealtec Block Heater has two models: HB-1 Block Heater (one block model) and HB-2 Block Heater (two blocks model). Both models can set heat temperature and timer control. The interchangeable aluminum blocks offer more applicable tube types. The digital control module can offer researchers more accurate temperature control for experiment requirements.




Individual functional descriptions ---


- a. On/Off: Power switch is positioned on backside of the instrument. When power is turned on, POWER indicator and LED display will light up. Operator shall connect the instrument to correct voltage source as indicated on the label of instrument.
- b. Power cord inlet: The inlet for the power cord is positioned on the backside of the Block Heater.


- c. : Start and Stop control button: Press Start/Stop button to start/ stop the running of the Block Heater after experiment condition has been set.

- d. : The LED display monitors can show all settings/ status value while setting or running. There are two LED display monitors on the Block Heater. One is to visualize temperature value (left side) and the other is to display setting/ remaining value of the timer (right side).

- e. : Up and Down arrow: Press up and down arrow to select the condition setting value. There are two sets of up and down arrows on Block Heater: Set temperature value (Left side)* and Set value of the timer (right side).

* **Note!** After start running, **do not** press Temperature Up and Down arrow. This will activate instruments OFFSET function. Offset function shall not be activated before running temperature is reached and stabilized. For further information about OFFSET function, please refer to OFFSET function section described below.

- f. : Indicator lights for Power, Heating and Set Temp: When power is turned on, both Power and Set Temp indicators lights up, inform operator to set desired temperature. When press Start/ Stop button to start heating, the Set Temp light will be off and Heating light on.

- g. : °F * and °C lights: Located on the left side of temperature LED display monitor. Block Heater's temperature can be displayed as Centigrade or Fahrenheit. Simultaneously press temperature up and down key to alternate Centigrade or Fahrenheit Temperature display.

* Note! Setting functions are not workable when displayed in Fahrenheit Temperature. Operator shall press back Centigrade Temperature display to activate setting functions.

- h. : Aluminum Blocks: Wealtec offers three types of blocks: 12 x 16mm test tubes; 20 x 1.5 ml microtubes and 30 x 0.5 ml microtubes

C. Installation of Block Heater

Package list:

Standard package

Item	Quantity
Block Heater main frame	1
Aluminum Block	1(HB-1) or 2(HB-2)
Power cable	1
Installation and Operation Instruction Manual	1

Installation:

1. Unpack the package and take the Block Heater out of the box.
2. Place the Block Heater on a stable, flat surface and proper environment for operation.
3. Make sure the power source has the correct voltage, current and frequency as indicated on the device.
4. Plug the power cable into instrument's inlet located at the backside of the Block Heater, then connect to the power source.

5. Turn on the power of Block Heater and check all functions as to Operation section described below. If the Block Heater has any functional defects, please contact your local representative or Wealtec worldwide service center for assistance.

D. Operation

1. Turn on the On/Off power switch, the Power light and LED monitor will light up. The temperature LED monitor will display default value of 30°C.
2. Press Temperature Up and Down-key to set the desired value. If the function is not workable, simultaneously press Up and Down-key to alternate display of Fahrenheit Temperature to Centigrade display.
3. Press Timer Up and Down-key to set running time. For continuously running without timer control, set the value to "0". Timer setting function will only be activated under "Stop" situations.
4. After all settings are completed, press Start/Stop button to initiate the Block Heater. By pressing start button, Heating indicator will light up and Set Temp light will be switched off. The Temperature LED monitor will flash and display the status of temperature. The flash will stop when set temperature is reached and stabilized. Timer LED display will start to count down after press start button.

● **Caution! Do not touch the hot top of the Block Heater while running**

5. To use the OFFSET function, operator can start operation after the temperature is reached and stabilized at set value while running. Please refer to OFFSET function section described below.
6. To stop the Block Heater manually, press Start/ Stop button.

E. Care and maintenance

Blocks shall not be touched while still hot. Operator shall not directly pour water/liquid/ice on the hot block/ block heater to reduce the temperature of the heating surface. It may cause damage to the instrument and fast-evaporated water/ liquid may be harmful to the operator.

After the block temperature cooled down, Block Heater shall be cleaned by moisten soft tissues after every usage to avoid all possible contaminations and damages to the instruments.

Aluminum blocks can be washed with water and mild detergent. Please avoid scrape the surface with hard brush, it may damage the specialized coating of aluminum block.

F. OFFSET function

- ➡ Important Notice: Every Block Heater has been properly calibrated by calibrated digital thermometer before product delivery. Operator is not recommended to use the OFFSET function and should only to be used if situation as below:

- A. Operator has a standard calibrated thermometer
- B. Operator wish to read the value same as their mercury thermometer to keep experimental conditions at the same standard.

*Note! General laboratory mercury thermometer is not calibrated from international standard thermometer. Some temperature variation may occur with different thermometer.

- ➡ If necessary to perform any adjustments in the field, please follow procedure as below:
 - A. Put a thermometer in the hole of the aluminum block and operate Block Heater in same way as the Operation section above to start the Block Heater.

- B. The OFFSET function is activated only under running condition and the temperature of block is reached and stabilized at the desired set value (Temperature LED display stops flashing at this moment).
- C. When press the Temperature Up or Down-key under the situation as described above, the Temperature LED display monitor will be off and Timer LED display will show 000.
- Note! The function display will be de-activated within 3 seconds if no further press on-Up or Down-key. Press Temperature Up or Down-key to activate the OFFSET function again.
- D. If desired value is lower then displayed value, press down arrow button of Temperature setting. If desired value is higher then displayed value, press up arrow button of Temperature setting. The maximum adjustable range is ± 100 (Equal to $\pm 10\text{ }^{\circ}\text{C}$) . In this function display, 1 degree equal to 0.1°C adjustment.

- E. For example, if operator's thermometer displays 98.5°C and Block Heater shows 100°C . Operator shall press down button of Temperature setting to -15 ($98.5-100= -1.5^{\circ}\text{C}$). After 3 seconds without any further action, display will automatically return to temperature and timer display.
 - F. Wait for 10 minutes to make sure the temperature of Block Heater becomes stable.
 - G. Check temperature with thermometer again and compare to display of the Block Heater. If the display value still differs from thermometer, repeat the OFFSET function setting again.
- ➡ **The OFFSET function will automatically return to default value after power turned off. Operator shall re-adjust to fit their requirement every time and for any reason power been turned/ switched or cut off.**

G. Order Information

Item #	Description
1092001	HB-1 Block Heater with an aluminum block for 12 x 16 mm test tubes, 120V
1092002	HB-1 Block Heater with an aluminum block for 12 x 16 mm test tubes, 230V
1092003	HB-1 Block Heater with an aluminum block for 20 x 1.5 ml microtubes, 120V
1092004	HB-1 Block Heater with an aluminum block for 20 x 1.5 ml microtubes, 230V
1092005	HB-1 Block Heater with an aluminum block for 30 x 0.5 ml microtubes, 120V
1092006	HB-1 Block Heater with an aluminum block for 30 x 0.5 ml microtubes, 230V
1092107	HB-2 Block Heater with two optional aluminum blocks*, 120V
1092108	HB-2 Block Heater with two optional aluminum blocks*, 230V

* Purchase order shall include the item numbers of optional blocks

Accessories:

1092101	Aluminum block for 12 x 16 mm test tubes
1092102	Aluminum block for 20 x 1.5 ml microtubes
1092103	Aluminum block for 30 x 0.5 ml microtubes
1092104	Block Extractor



WEALTEC CORP.

1885 Meadowvale Way
Sparks, NV 89431

TEL: +1-775 351 2066

FAX: +1-775 351 2077

E-mail : sales@wealtec.com
support@wealtec.com

Web site: www.wealtec.com

Warranty Card

THANK YOU FOR ORDERING A WEALTEC PRODUCT.

Wealtec Corp. warrants all Wealtec instruments to be free from defects in design, workmanship, and material under normal use for a period of one year from the date of initial shipment.

This warranty covers all parts and components of the instrument except those normally requiring frequent replacement, such as tubing, gasket, O-rings, etc. Wealtec will not be liable for any personal injury, bodily injury, misuse, improper maintenance, negligence or accident.

This warranty is in lieu of all other warranties, expressed or implied, but not limited to, the implied warranties of merchantability or fitness for a particular purpose.

PLEASE KEEP THE WARRANTY CARD FOR FUTURE USE.

Instrument Model :

Item Number :

Serial Number : EDW07J1005-V2

Initial Shipping:

Packager
Identification

Wealtec Corp.

1885 Meadowvale Way Sparks, NV 89431 Tel:+1-775-351-2066 Fax:+1-775-351-2077
<http://www.wealtec.com> E-mail: support@wealtec.com